



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE

United States Patent and Trademark Office

Address: COMMISSIONER FOR PATENTS

P.O. Box 1450

Alexandria, Virginia 22313-1450

www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/516,432	07/29/2005	Manfred Dick	134.1031	5126
7278	7590	07/08/2009		
DARBY & DARBY P.C. P.O. BOX 770 Church Street Station New York, NY 10008-0770			EXAMINER FARAH, AHMED M	
			ART UNIT	PAPER NUMBER
			3769	
			MAIL DATE	DELIVERY MODE
			07/08/2009 PAPER	

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary

Application No.

10/516,432

Applicant(s)

DICK ET AL.

Examiner

Ahmed M. Farah

Art Unit

3769

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --
Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 11 March 2009.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 14-28 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 14-28 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- 1) ☒ Notice of References Cited (PTO-892)
- 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- 3) ☐ Information Disclosure Statement(s) (PTO/SG/US)
Paper No(s)/Mail Date _____
- 4) ☐ Interview Summary (PTO-413)
Paper No(s)/Mail Date _____
- 5) ☐ Notice of Informal Patent Application
- 6) ☐ Other: _____

DETAILED ACTION

Claim Rejections - 35 USC § 112

The following is a quotation of the second paragraph of 35 U.S.C. 112:

The specification shall conclude with one or more claims particularly pointing out and distinctly claiming the subject matter which the applicant regards as his invention.

Claims 14-24 are rejected under 35 U.S.C. 112, second paragraph, as being indefinite for failing to particularly point out and distinctly claim the subject matter which applicant regards as the invention.

Claim 14 is rejected under 35 U.S.C. 112, second paragraph, as being incomplete for omitting essential elements, such omission amounting to a gap between the elements. **See MPEP § 2172.01.** The omitted element is an 'ablation device' being controlled according to the determined optic and geometrical data of the eye as claimed.

Note: the preamble of independent claim 14 recites method for controlling an ablation device (laser source). In the body of the claim, the applicant recites the step of "controlling the [ablation] device." However, the applicant fails to positively recite the ablation device in the body of the claim.

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claim 14 is rejected under 35 U.S.C. 102(b) as being anticipated by Ruiz US Patent No. 6,129,722.

Ruiz discloses ophthalmic apparatus and method for correcting vision disorders of the eye, the method comprising:

determining optic and geometrical data of the eye (topographically mapping in detail the irregularities and surface deviations of the cornea);

performing a graphical simulation of the ablation in the form of a graphic visualization (providing an interface system, which receives the optic and geometrical data and allows the surgeon to simulate and visualize surgical alternatives); and

controlling a surgical laser system for ablation of the eye according to the graphic data as claimed (receiving and manipulating the obtained topographical data and providing control command to a surgical laser). See the abstract; col. 14, lines 44-48; and claims 29 and 34.

Claims 14-28 are again rejected under 35 U.S.C. 102(e) as being anticipated by Cox et al. US Patent No. 7,130,835.

Cox et al. disclose apparatus and methods of use for providing refractive corrections of a patient's eye, the apparatus comprising: a treatment laser for generating ablative laser beams; a diagnostic system such as wavefront sensor, aberrometer, pachymeter, topographer, and tonometer; and an electronic data

processing and control system for controlling refractive correction of the eye as claimed (see Figures 1-4).

With respect to the limitation "controlling the device for the ablation" as recited in amended claim 14, Cox et al. further teach a method of controlling a surgical laser based on determined topographic data of the eye (see the abstract and Figs. 1-4).

Conclusion

The prior art made of record and not relied upon is considered pertinent to applicant's disclosure.

US patent No. 7,111,938 to Andino et al. discloses a lens design and manufacturing system wherein corneal topography of the eye is obtained, displayed in a computer system, and retinal image quality is simulated and displayed in the same computer system as graphical representation.

Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the

shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Ahmed M. Farah whose telephone number is (571) 272-4765. The examiner can normally be reached on Mon, Tue, Thur and Fri between 9:30 AM 7:30 PM.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Johnson Henry can be reached on (571) 272-4768. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/Ahmed M Farah/
Primary Examiner, Art Unit 3769

July 4, 2009.